



EC DECLARATION OF CONFORMITY

Manufacturer: ELLA-CS, s.r.o.

Milady Horákové 504/45 Třebeš
500 06 Hradec Králové 6
Czech Republic

Product: Extractor for SX-ELLA Stents Esophageal (ELLA Extractor)
(*Non-active invasive medical device*)

Classification: a) class: II a
b) rule: 5

Conformity assessment route: Annex II to Council Directive 93/42/EEC amended by Directive 2007/47/EC

We herewith declare that the above-mentioned product meets the provisions of the Council Directive 93/42/EEC and Directive 2007/47/EC for medical devices. All supporting documentation is retained under the premises of the manufacturer.

Standards applied:

- EN ISO 13485:2012 Medical devices – Quality management systems – Requirements for regulatory purposes
- EN ISO 14630:2012 Non-active surgical implants – General requirements
- EN ISO 14971:2012 Medical devices – Application of risk management to medical devices
- EN ISO 10993-1:2009/AC:2010 Biological evaluation of medical devices – Part 1: Evaluation and testing
- EN ISO 10993-5:2009 Biological evaluation of medical devices – Part 5: Tests for in vitro cytotoxicity
- EN ISO 10993-7:2008/AC:2009 Biological evaluation of medical devices – Part 7: Ethylene oxide sterilization residuals
- EN ISO 14644-1:2015 Cleanrooms and associated controlled environments – Part 1: Classification of air cleanliness
- EN ISO 14644-2:2015 Cleanrooms and associated controlled environments – Part 2: Specifications for testing and monitoring to prove continued compliance with ISO 14644-1
- EN ISO 11135:2014 Sterilization of health care products - Ethylene oxide - Requirements for the development, validation and routine control of a sterilization process for medical devices
- EN ISO 11737-1:2006 Sterilization of medical devices – Microbiological methods – Part 1: Determination of a population of microorganisms on products
- EN 556-1:2001 Sterilization of medical devices – Requirements for medical devices to be designated “STERILE” – Part 1: Requirements for terminally sterilized medical devices



- EN ISO 11138-1:2006 Sterilization of health care products – Biological indicators – Part 1: General requirements
- EN ISO 11138-2:2009 Sterilization of health care products – Biological indicators – Part 2: Biological indicators for ethylene oxide sterilization processes
- EN ISO 11607-1:2009/A1:2014 Packaging for terminally sterilized medical devices – Part 1: Requirements for materials, sterile barrier systems and packaging systems
- EN 1041+A1:2013 Information supplied by the manufacturer of medical devices
- EN ISO 15223-1:2016 Medical devices – Symbols to be used with medical device labels, labeling and information to be supplied – Part 1: General requirements
- MEDDEV 2.12/1 in the current version: Guidelines on a medical devices vigilance system

Notified Body: Electrotechnical Testing Institute
Pod lisem 129
Prague 8 – Troja
Czech Republic
Notified body No. 1014

(EC) Certificate: MED 170033 – issued on June 29, 2017; valid until June 28, 2022

CE marked since: Date: 2007-01
Lot Number: 150060107 (2 pieces)

Place of issue: Hradec Králové, Czech Republic
Date of issue: July 17, 2017

ELLA-CS, s.r.o.
PRODUCTION OF MEDICAL DEVICES
Milady Horákové 504/45, Trebeš
500 06 Hradec Králové, Czech Republic
tel.: +420 495 279 111, fax: +420 495 265 655
VAT No.: CZ27507785

Assoc. Prof. Karel Volenec, Ph.D.
Executive Head of ELLA-CS, s.r.o.